



Member of the FM Global Group

FM Approvals  
1151 Boston Providence Turnpike  
P.O. Box 9102 Norwood, MA 02062 USA  
T: 781 762 4300 F: 781-762-9375 www.fmapprovals.com

# CERTIFICATE OF COMPLIANCE

## HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT

This certificate is issued for the following equipment:

**A35, B35, BBS, BLC, BLF, C2, C3H2, C3H3, C6A, C16, DEBA, H35, HLC, JRT, LCB, MBB, PLC, PLX, PW12, PW16, PW2F-2, PW2FC3, PW2G, PW6, PW10, PW12, PW15, PW16, PWA, PWS, RSC, RTN, S35, SB2, SB3, SBF, SP4, SS4, THC, TLC, U1A, U2A, U2B, Z6FD1, Z6FC3. Load Cells**

IS / I,II,III / 1 / ABCDEFG — 3P14667 and /or 3P14595 ; Entity\*; NI / I / 2 / ABCD; S / II,III / 2 / FG

\*Refer to Dwg. 3P14667 and /or 3P14595 for entity parameters.

### Equipment Ratings:

Intrinsically Safe for use in Class I, II and III, Division 1, Groups A, B, C, D, E, F and G: Temperature class T4 when installed in accordance with Entity requirements and control drawing 3P14667 and /or 3P14595; Nonincendive for use in Class I, Division 2, Groups A, B, C and D; Suitable for use in Class II and III, Division 2, Groups F and G Hazardous (Classified) Locations

### FM Approved for:

HBM Inc.  
19 Bartlett St.  
Marlboro, MA 01752

This certifies that the equipment described has been found to comply with the following Approval Standards and other documents:

Class 3600	1998
Class 3610	2010
Class 3810	1989
Class 3611	2004

Original Project ID: 0T5A2.AX

Approval Granted: September 20, 1991

Subsequent Revision Reports / Date Approval Amended

Report Number	Date	Report Number	Date
1Q1A6.AX	September 20, 1991		
3W1A1.AX	August 7, 1992		
3W6A8.AX	September 23, 1992		
1Z7A7.AX	August 24, 1995		
1B2A6.AX	May 30, 1996		
3005091	July 20, 1999		
3006081	January 7, 2000		
3007883	June 13, 2000		
3009841	December 12, 2000		
040719	August 12, 2004		
050815	September 9, 2005		
060315	June 9, 2006		
110324	March 29, 2011		
3042054	February 1, 2012		

FM Approvals LLC



J. E. Marquedant  
Group Manager, Electrical

1 February 2012  
Date